DEVICE SPECIFICATIONS

Bus-Powered Multifunction DAQ USB Device

The following specifications are typical at 25 °C, unless otherwise noted. For more information about the NI USB-6009, refer to the *NI USB-6008/6009 User Guide* available from *ni.com/manuals*.

Analog Input

Analog inputs	
Differential	4
Single-ended	8, software-selectable
Input resolution	
Differential	14 bits
Single-ended	13 bits
Maximum sample rate (aggregate)	48 kS/s, system dependent
Converter type	Successive approximation
AI FIFO	512 bytes
Timing resolution	41.67 ns (24 MHz timebase)
Timing accuracy	100 ppm of actual sample rate
Input range	
Differential	±20 V ¹ , ±10 V, ±5 V, ±4 V, ±2.5 V, ±2 V, ±1.25 V, ±1 V
Single-ended	±10 V
Working voltage	±10 V
Input impedance	144 kΩ



¹ ± 20 V means that $|AI+ - (AI-)| \le 20$ V. However, AI+ and AI- must both be within ± 10 V of GND. Refer to the *Taking Differential Measurements* section of the *NI USB-6008/6009 User Guide* for more information.

Overvoltage protection	±35 V
Trigger source	Software or external digital trigger
System noise ²	
Differential	
± 20 V range	5 mV _{rms}
±1 V range	0.5 mV _{rms}
Single-ended, ± 10 V range	5 mV _{rms}

Table 1. Absolute Accuracy at Full Scale, Differential

Range (V)	Typical at 25 °C (mV)	Maximum over Temperature (mV)
±20	14.7	138
±10	7.73	84.8
±5	4.28	58.4
±4	3.59	53.1
±2.5	2.56	45.1
±2	2.21	42.5
±1.25	1.70	38.9
±1	1.53	37.5



Note Input voltages may not exceed the working voltage range.

Table 2. Absolute Accuracy at Full Scale, Single-Ended

Range (V)	Typical at 25 °C (mV)	Maximum over Temperature (mV)
±10	14.7	138

Analog Output

Analog outputs	2
Output resolution	12 bits
Maximum update rate	150 Hz, software-timed

² System noise measured at maximum sample rate.

Output range	0 V to +5 V
Output impedance	50 Ω
Output current drive	5 mA
Power-on state	0 V
Slew rate	1 V/µs
Short circuit current	50 mA
Absolute accuracy (no load)	
Typical	7 mV
Maximum at full scale	36.4 mV

Digital I/O

8 lines
4 lines
Each channel individually programmable as input or output
Each channel individually programmable as open collector or active drive
TTL, LVTTL, CMOS
-0.5 V to 5.8 V with respect to GND
$4.7 \text{ k}\Omega$ to 5 V
Input

Table 3. Digital Logic Levels

Level	Minimum	Maximum
Input low voltage	-0.3 V	0.8 V
Input high voltage	2.0 V	5.8 V
Input leakage current		50 µA
Output low voltage (I = 8.5 mA)		0.8 V
Output high voltage, active drive (I = -8.5 mA)	2.0 V	3.5 V

³ This document uses NI-DAQmx naming conventions. Open-drain is called open collector and push-pull is called active drive.

Level	Minimum	Maximum
Output high voltage, open collector (I = -0.6 mA, nominal)	2.0 V	5.0 V
Output high voltage, open collector (I = -8.5 mA, with external pull-up resistor)	2.0 V	_

Table 3. Digital Logic Levels (Continued)

External Voltage

+5 V output (200 mA maximum)	
Minimum	+4.85 V
Typical	+5 V
+2.5 V output (1 mA maximum)	+2.5 V
+2.5 V accuracy	0.25% maximum
Reference temperature drift	50 ppm/°C maximum

Event Counter

Number of counters	1
Resolution	32 bits
Counter measurements	Edge counting (falling-edge)
Counter direction	Count up
Pull-up resistor	$4.7 \text{ k}\Omega$ to 5 V
Maximum input frequency	5 MHz
Minimum high pulse width	100 ns
Minimum low pulse width	100 ns
Input high voltage	2.0 V
Input low voltage	0.8 V

Bus Interface

USB specification

USB 2.0 full-speed (12 Mb/s)


Power Requirements

80 mA
500 mA
300 µA
500 μΑ

Physical Characteristics

Dimensions	
Without connectors	63.5 mm × 85.1 mm × 23.2 mm (2.50 in. × 3.35 in. × 0.91 in.)
With connectors	81.8 mm × 85.1 mm × 23.2 mm (3.22 in. × 3.35 in. × 0.91 in.)
Weight	
Without connectors	54 g (1.9 oz)
With connectors	84 g (3 oz)
I/O connectors	USB series B receptacle, (2) 16-position screw terminal plugs
Screw-terminal wiring	16 AWG to 28 AWG
Torque for screw terminals	$0.22 \text{ N} \cdot \text{m}$ to $0.25 \text{ N} \cdot \text{m}$ (2.0 lb \cdot in. to 2.2 lb \cdot in.)

If you need to clean the module, wipe it with a dry towel.



Safety Voltages

Connect only voltages that are within these limits.

Channel-to-GND

±30 V max, Measurement Category I

Measurement Category I is for measurements performed on circuits not directly connected to the electrical distribution system referred to as MAINS voltage. MAINS is a hazardous live electrical supply system that powers equipment. This category is for measurements of voltages from specially protected secondary circuits. Such voltage measurements include signal levels, special equipment, limited-energy parts of equipment, circuits powered by regulated low-voltage sources, and electronics.



Caution Do not use this module for connection to signals or for measurements within Measurement Categories II, III, or IV.

Environmental

Temperature (IEC 60068-2-1 and IE	C 60068-2-2)
Operating	0 °C to 55 °C
Storage	-40 °C to 85 °C
Humidity (IEC 60068-2-56)	
Operating	5% RH to 95% RH, noncondensing
Storage	5% RH to 90% RH, noncondensing
Pollution Degree (IEC 60664)	2
Maximum altitude	2,000 m

Indoor use only.

Safety

This product is designed to meet the requirements of the following electrical equipment safety standards for measurement, control, and laboratory use:

- IEC 61010-1, EN 61010-1
- UL 61010-1, CSA 61010-1



Note For UL and other safety certifications, refer to the product label or the *Online Product Certification* section.

Electromagnetic Compatibility

This product meets the requirements of the following EMC standards for electrical equipment for measurement, control, and laboratory use:

- EN 61326-1 (IEC 61326-1): Class A emissions; Basic immunity
- EN 55011 (CISPR 11): Group 1, Class A emissions
- EN 55022 (CISPR 22): Class A emissions
- EN 55024 (CISPR 24): Immunity
- AS/NZS CISPR 11: Group 1, Class A emissions
- AS/NZS CISPR 22: Class A emissions
- FCC 47 CFR Part 15B: Class A emissions
- ICES-001: Class A emissions



Note In the United States (per FCC 47 CFR), Class A equipment is intended for use in commercial, light-industrial, and heavy-industrial locations. In Europe, Canada, Australia and New Zealand (per CISPR 11) Class A equipment is intended for use only in heavy-industrial locations.



Note Group 1 equipment (per CISPR 11) is any industrial, scientific, or medical equipment that does not intentionally generate radio frequency energy for the treatment of material or inspection/analysis purposes.



Note For EMC declarations and certifications, and additional information, refer to the *Online Product Certification* section.

CE Compliance $C \in$

This product meets the essential requirements of applicable European Directives, as follows:

- 2014/35/EU; Low-Voltage Directive (safety)
- 2014/30/EU; Electromagnetic Compatibility Directive (EMC)

Online Product Certification

Refer to the product Declaration of Conformity (DoC) for additional regulatory compliance information. To obtain product certifications and the DoC for this product, visit *ni.com/ certification*, search by model number or product line, and click the appropriate link in the Certification column.

Environmental Management

NI is committed to designing and manufacturing products in an environmentally responsible manner. NI recognizes that eliminating certain hazardous substances from our products is beneficial to the environment and to NI customers.

For additional environmental information, refer to the *Minimize Our Environmental Impact* web page at *ni.com/environment*. This page contains the environmental regulations and directives with which NI complies, as well as other environmental information not included in this document.


Waste Electrical and Electronic Equipment (WEEE)

EU Customers This symbol indicates that waste products should be disposed of separately from municipal household waste according to WEEE Directive 2002/96/EC of the European Parliament and the Council on waste electrical and electronic equipment (WEEE). All products at the end of their life cycle must be sent to a WEEE collection and recycling center. Proper WEEE disposal reduces environmental impact and the risk to human health due to potentially hazardous substances used in such equipment. Your cooperation in proper WEEE disposal will contribute to the effective usage of natural resources. For information about the available collection and recycling scheme in a particular country, go to *ni.com/environment/weee*.

电子信息产品污染控制管理办法(中国 RoHS)

中国客户 National Instruments 符合中国电子信息产品中限制使用某些有害物质指令(RoHS)。关于 National Instruments 中国 RoHS 合规性信息,请登录 ni.com/environment/rohs_china。(For information about China RoHS compliance, go to ni.com/environment/rohs_china.)

Device Pinout



Refer to the *NI Trademarks and Logo Guidelines* at ni.com/trademarks for information on National Instruments trademarks. Other product and company names mentioned herein are trademarks or trade names of their respective companies. For patents covering National Instruments products/technology, refer to the appropriate location: **Help»Patents** in your software, the patents.txt file on your media, or the *National Instruments Patent Notice* at ni.com/patents. You can find information about end-user license agreements (EULAs) and third-party legal notices in the readme file for your NI product. Refer to the *Export Compliance Information* at ni.com/legal/export-compliance for the National Instruments global trade compliance policy and how to obtain relevant HTS codes, ECCNs, and other import/export data. NI MAKES NO EXPRESS OR IMPLIED WARRANTIES AS TO THE ACCURACY OF THE INFORMATION CONTAINED HEREIN AND SHALL NOT BE LIABLE FOR ANY ERRORS. U.S. Government Customers: The data contained in this manual was developed at private expense and is subject to the applicable limited rights and restricted data rights as set forth in FAR 52.227-14, DFAR 252.227-7014, and DFAR 252.227-7015.

© 2004-2015 National Instruments. All rights reserved.